

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

MAY 30 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130309</u>	
Name <u>PATRICK NELSON</u>		ID# <u>4389</u>	Date <u>5-29-09</u>
A	Agency <u>AST</u>	Phone # <u>373-8337</u>	
Instrument Location <u>MATSU WEST, WASILLA</u>			
Alco S/N <u>70729</u>		Target Value <u>.082</u>	High Pressure <u>900</u>
B	Alco Test Values		
	<u>.082</u> 1 st Alco	<u>.082</u> 2 nd Alco	
Signature <u>[Signature]</u>		<u>CB</u> <u>6/6/09</u>	
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 4th day of June 2009.

[Signature]
Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130309 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

MAY 29, 2009

OPERATOR'S NAME: COB 62109
NELSON PATRICK
OPERATOR'S NUMBER: 4389
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF CALIBRATION
O.L. #: 123456
DEPT/AGENCY: PLMB
CASE/REPORT: 12-3456
TEST TYPE: U
ALCO TARGET VALUE: .002 ✓
ALCO S/N: 70729 ✓

— BREATH ANALYSIS —

.002 ADJUSTED FOR 29.73 in ✓
ALCO TARGET .001 12:25
BLANK TEST .000 12:26
INTERNAL STANDARD VERIFIED 12:26
ALCO TV 29.73 in .002 ✓ 12:26
BLANK TEST .000 12:27
SUBJECT SAMPLE COB 62109 .000 12:28
BLANK TEST .000 12:29
ALCO TV 29.73 in .002 ✓ 12:29
BLANK TEST .000 12:30

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

MAY 29, 2009 ✓
TIME 12:22

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.73 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~
printtest()